

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
 Data File : VX022282.D
 Acq On : 29 May 2021 03:51
 Operator : JC/MD
 Sample : M2532-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-N-(MH-10.29)

Quant Time: May 29 05:00:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	71856	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	153137	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	135487	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	55997	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.965	65	62166	56.08	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	112.16%	
35) Dibromofluoromethane	5.398	113	45436	46.79	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	93.58%	
50) Toluene-d8	8.653	98	187859	50.22	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	100.44%	
62) 4-Bromofluorobenzene	11.086	95	66953	48.42	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	96.84%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	423	0.42	ug/l #	71
16) Acetone	2.386	43	2379	4.93	ug/l	96
27) cis-1,2-Dichloroethene	4.501	96	2727	2.50	ug/l	91
44) Trichloroethene	7.129	130	6279	5.80	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052921\
Data File : VX022282.D
Acq On : 29 May 2021 03:51
Operator : JC/MD
Sample : M2532-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-N-(MH-10.29)

Quant Time: May 29 05:00:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

