

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX052924\  
 Data File : VX041672.D  
 Acq On : 29 May 2024 14:29  
 Operator : JC/MD  
 Sample : VX0529WBL01  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 14 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 VX0529WBL01

Quant Time: May 30 07:41:49 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X052924W.M  
 Quant Title : SW846 8260  
 QLast Update : Thu May 30 07:36:24 2024  
 Response via : Initial Calibration

| Compound                    | R.T.   | QIon  | Response | Conc     | Units | Dev(Min) |
|-----------------------------|--------|-------|----------|----------|-------|----------|
| -----                       |        |       |          |          |       |          |
| Internal Standards          |        |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.550  | 168   | 141338   | 50.000   | ug/l  | # 0.00   |
| 34) 1,4-Difluorobenzene     | 6.757  | 114   | 214425   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055 | 117   | 206342   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024 | 152   | 99182    | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |        |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.952  | 65    | 81120    | 46.796   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 78 - 117 | Recovery | =     | 93.600%  |
| 35) Dibromofluoromethane    | 5.385  | 113   | 67252    | 45.363   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 75 - 124 | Recovery | =     | 90.720%  |
| 50) Toluene-d8              | 8.647  | 98    | 247554   | 47.789   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 92 - 112 | Recovery | =     | 95.580%  |
| 62) 4-Bromofluorobenzene    | 11.079 | 95    | 91199    | 47.701   | ug/l  | 0.00     |
| Spiked Amount               | 50.000 | Range | 83 - 123 | Recovery | =     | 95.400%  |

|                  |        |
|------------------|--------|
| Target Compounds | Qvalue |
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX052924\  
Data File : VX041672.D  
Acq On : 29 May 2024 14:29  
Operator : JC/MD  
Sample : VX0529WBL01  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 14 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
VX0529WBL01

Quant Time: May 30 07:41:49 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X052924W.M  
Quant Title : SW846 8260  
QLast Update : Thu May 30 07:36:24 2024  
Response via : Initial Calibration

