

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052924\
 Data File : VX041676.D
 Acq On : 29 May 2024 16:06
 Operator : JC/MD
 Sample : PB161143ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161143ZHE#02

Quant Time: May 30 07:44:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143446	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	219037	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	218036	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	107786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	82945	47.146	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.300%
35) Dibromofluoromethane	5.385	113	69189	45.687	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.380%
50) Toluene-d8	8.647	98	260061	49.147	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.300%
62) 4-Bromofluorobenzene	11.079	95	98698	50.537	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.080%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	25570	30.492	ug/l	92
18) Methyl Acetate	2.709	43	2692	1.603	ug/l	99
20) Methylene Chloride	2.788	84	10977	6.972	ug/l	91
43) Isopropyl Acetate	6.342	43	118477	33.909	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052924\
Data File : VX041676.D
Acq On : 29 May 2024 16:06
Operator : JC/MD
Sample : PB161143ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161143ZHE#02

Quant Time: May 30 07:44:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

