

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052924\
 Data File : VX041678.D
 Acq On : 29 May 2024 16:53
 Operator : JC/MD
 Sample : PB161143ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161143ZHE#04

Quant Time: May 30 07:45:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	151874	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	233315	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	231105	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	86093	46.220	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	92.440%
35) Dibromofluoromethane	5.385	113	72646	45.034	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.060%
50) Toluene-d8	8.647	98	278708	49.447	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.900%
62) 4-Bromofluorobenzene	11.079	95	102646	49.342	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.680%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	66838	79.686	ug/l	96
18) Methyl Acetate	2.709	43	3211	1.805	ug/l	94
20) Methylene Chloride	2.788	84	9917	5.949	ug/l	# 91
25) 2-Butanone	4.574	43	10128	8.749	ug/l	96
43) Isopropyl Acetate	6.342	43	125032	33.595	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052924\
Data File : VX041678.D
Acq On : 29 May 2024 16:53
Operator : JC/MD
Sample : PB161143ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161143ZHE#04

Quant Time: May 30 07:45:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

