

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052924\
 Data File : VX041679.D
 Acq On : 29 May 2024 17:16
 Operator : JC/MD
 Sample : PB161143ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB161143ZHE#05

Quant Time: May 30 07:45:56 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
 Quant Title : SW846 8260
 QLast Update : Thu May 30 07:36:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	148380	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	225159	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	224824	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	113968	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	83317	45.783	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	91.560%
35) Dibromofluoromethane	5.385	113	69312	44.524	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.040%
50) Toluene-d8	8.647	98	267168	49.117	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.240%
62) 4-Bromofluorobenzene	11.079	95	102170	50.892	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.780%
Target Compounds						
						Qvalue
16) Acetone	2.398	43	67539	82.538	ug/l	94
18) Methyl Acetate	2.715	43	5415	3.116	ug/l	99
20) Methylene Chloride	2.788	84	9185	5.640	ug/l	94
25) 2-Butanone	4.592	43	9550	8.444	ug/l	100
43) Isopropyl Acetate	6.348	43	115432	32.139	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052924\
Data File : VX041679.D
Acq On : 29 May 2024 17:16
Operator : JC/MD
Sample : PB161143ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB161143ZHE#05

Quant Time: May 30 07:45:56 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Quant Title : SW846 8260
QLast Update : Thu May 30 07:36:24 2024
Response via : Initial Calibration

