



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052924\
Data File : VX041661.D
Acq On : 29 May 2024 08:43
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052924W.M
Title : SW846 8260
Last Update : Thu May 30 07:36:24 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.7	7753	PASS
75	95	30	60	52.5	20709	PASS
95	95	100	100	100.0	39435	PASS
96	95	5	9	6.9	2709	PASS
173	174	0.00	2	0.8	324	PASS
174	95	50	100	96.8	38184	PASS
175	174	5	9	7.4	2813	PASS
176	174	95	101	95.1	36331	PASS
177	176	5	9	6.8	2469	PASS