

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052925\
 Data File : VX046404.D
 Acq On : 29 May 2025 17:07
 Operator : JC/MD
 Sample : Q2155-04 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 444

Quant Time: May 30 01:33:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	63188	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	125445	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	119851	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51885	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63916	54.257	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.520%	
35) Dibromofluoromethane	5.379	113	45393	50.250	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.500%	
50) Toluene-d8	8.647	98	159553	51.031	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.060%	
62) 4-Bromofluorobenzene	11.079	95	62048	51.737	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.480%	
Target Compounds						
16) Acetone	2.380	43	5093	10.783	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX052925\
Data File : VX046404.D
Acq On : 29 May 2025 17:07
Operator : JC/MD
Sample : Q2155-04 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
444

Quant Time: May 30 01:33:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

