


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX053019\
Data File : VX009950.D
Acq On : 30 May 2019 10:32
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Title : SW846 8260
Last Update : Wed May 29 04:00:56 2019



AutoFind: Scans 1636, 1637, 1638; Background Corrected with Scan 1630

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	22.2	6276	PASS
75	95	30	60	51.5	14547	PASS
95	95	100	100	100.0	28224	PASS
96	95	5	9	6.7	1887	PASS
173	174	0.00	2	0.8	165	PASS
174	95	50	100	77.4	21851	PASS
175	174	5	9	7.9	1721	PASS
176	174	95	101	96.9	21177	PASS
177	176	5	9	6.7	1410	PASS