

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
 Data File : VX035900.D
 Acq On : 30 May 2023 12:54
 Operator : JC/MD
 Sample : 02945-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TWP03

Quant Time: May 30 16:31:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 29 07:45:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	211078	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	346669	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	309511	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	143708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	162823	43.408	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	86.820%		
35) Dibromofluoromethane	5.385	113	108931	45.745	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.480%		
50) Toluene-d8	8.647	98	414447	47.424	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	94.840%		
62) 4-Bromofluorobenzene	11.079	95	160879	46.724	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.440%		
Target Compounds						
16) Acetone	2.386	43	2512	1.659	ug/l	98
19) Methyl tert-butyl Ether	3.117	73	3612	0.404	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
Data File : VX035900.D
Acq On : 30 May 2023 12:54
Operator : JC/MD
Sample : 02945-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TWP03

Quant Time: May 30 16:31:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
Quant Title : SW846 8260
QLast Update : Mon May 29 07:45:36 2023
Response via : Initial Calibration

