

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
 Data File : VX035910.D
 Acq On : 30 May 2023 16:47
 Operator : JC/MD
 Sample : 02852-02ME 5X
 Misc : 6.60g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CB-32-G3-SO-14.5-051823ME

Quant Time: May 31 05:48:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 29 07:45:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	228263	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	377252	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	332811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158840	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	167709	41.344	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	82.680%
35) Dibromofluoromethane	5.385	113	113968	43.980	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.960%
50) Toluene-d8	8.647	98	440608	46.330	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.660%
62) 4-Bromofluorobenzene	11.079	95	177435	47.355	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.700%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.489	96	21592	6.718	ug/l	89
44) Trichloroethene	7.123	130	104066	37.905	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
Data File : VX035910.D
Acq On : 30 May 2023 16:47
Operator : JC/MD
Sample : 02852-02ME 5X
Misc : 6.60g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CB-32-G3-SO-14.5-051823ME

Quant Time: May 31 05:48:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
Quant Title : SW846 8260
QLast Update : Mon May 29 07:45:36 2023
Response via : Initial Calibration

