

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
 Data File : VX035912.D
 Acq On : 30 May 2023 17:34
 Operator : JC/MD
 Sample : 02902-01DL 50X
 Misc : 1.15g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FMC-0096DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 05/31/2023
 Supervised By :Mahesh Dadoda 05/31/2023

Quant Time: May 30 19:01:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 29 07:45:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	226441	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	373828	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	341417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	164887	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	166472	41.370	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	82.740%		
35) Dibromofluoromethane	5.385	113	114079	44.426	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	88.860%		
50) Toluene-d8	8.647	98	446155	47.343	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	94.680%		
62) 4-Bromofluorobenzene	11.079	95	178753	48.143	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	96.280%		
Target Compounds						Qvalue
16) Acetone	2.386	43	144435	88.907	ug/l	99
31) Cyclohexane	5.550	56	10114m	2.116	ug/l	
39) Methylcyclohexane	7.379	83	2686	0.604	ug/l	92
52) Toluene	8.720	92	127936	18.652	ug/l	96
67) Ethyl Benzene	10.195	91	12237	0.907	ug/l	94
68) m/p-Xylenes	10.299	106	16691	3.345	ug/l	96
69) o-Xylene	10.640	106	7554	1.524	ug/l	98
73) Isopropylbenzene	10.963	105	2707	0.202	ug/l	96
78) n-propylbenzene	11.305	91	7750	0.497	ug/l	100
80) 1,3,5-Trimethylbenzene	11.451	105	9922	0.871	ug/l	96
84) 1,2,4-Trimethylbenzene	11.750	105	38227	3.350	ug/l	99
89) n-Butylbenzene	12.329	91	4889	0.487	ug/l #	53
95) Naphthalene	13.774	128	12119	1.043	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
Data File : VX035912.D
Acq On : 30 May 2023 17:34
Operator : JC/MD
Sample : 02902-01DL 50X
Misc : 1.15g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FMC-0096DL

Quant Time: May 30 19:01:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
Quant Title : SW846 8260
QLast Update : Mon May 29 07:45:36 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 05/31/2023
Supervised By :Mahesh Dadoda 05/31/2023

