

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
 Data File : VX035914.D
 Acq On : 30 May 2023 18:21
 Operator : JC/MD
 Sample : 02924-02RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001-002-003-2RE

Quant Time: May 31 05:48:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 29 07:45:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	226998	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	371601	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	327213	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	151880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	145251	36.008	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	72.020%#
35) Dibromofluoromethane	5.385	113	109551	42.918	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	85.840%
50) Toluene-d8	8.647	98	430085	45.911	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.820%#
62) 4-Bromofluorobenzene	11.079	95	162382	43.996	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.000%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	12282	7.542	ug/l	98
20) Methylene Chloride	2.788	84	4170	0.587	ug/l #	93
30) Chloroform	5.093	83	9001	1.608	ug/l	97
47) Bromodichloromethane	7.824	83	5038	1.375	ug/l	97
60) Dibromochloromethane	9.525	129	2899	1.172	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
Data File : VX035914.D
Acq On : 30 May 2023 18:21
Operator : JC/MD
Sample : 02924-02RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001-002-003-2RE

Quant Time: May 31 05:48:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
Quant Title : SW846 8260
QLast Update : Mon May 29 07:45:36 2023
Response via : Initial Calibration

