

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
 Data File : VX035915.D
 Acq On : 30 May 2023 18:44
 Operator : JC/MD
 Sample : 02924-03RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001-002-003-3RE

Quant Time: May 31 05:48:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 29 07:45:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	229340	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	372067	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	327885	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	152870	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	144178	35.377	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	70.760%#
35) Dibromofluoromethane	5.385	113	110523	43.245	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.500%
50) Toluene-d8	8.647	98	431708	46.027	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.060%
62) 4-Bromofluorobenzene	11.079	95	161997	43.837	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	87.680%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	12343	7.502	ug/l	94
20) Methylene Chloride	2.788	84	4639	0.732	ug/l	97
30) Chloroform	5.099	83	9473	1.675	ug/l	96
47) Bromodichloromethane	7.824	83	4931	1.344	ug/l #	94
60) Dibromochloromethane	9.525	129	2086	0.842	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
Data File : VX035915.D
Acq On : 30 May 2023 18:44
Operator : JC/MD
Sample : 02924-03RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001-002-003-3RE

Quant Time: May 31 05:48:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
Quant Title : SW846 8260
QLast Update : Mon May 29 07:45:36 2023
Response via : Initial Calibration

