

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
 Data File : VX035916.D
 Acq On : 30 May 2023 19:08
 Operator : JC/MD
 Sample : 02924-01RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001-002-003-1RE

Quant Time: May 31 05:49:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
 Quant Title : SW846 8260
 QLast Update : Mon May 29 07:45:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	223641	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	367513	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	331732	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	158868	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	145060	36.500	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	73.000%#
35) Dibromofluoromethane	5.385	113	110520	43.780	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.560%
50) Toluene-d8	8.647	98	432491	46.681	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.360%
62) 4-Bromofluorobenzene	11.079	95	165340	45.296	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.600%
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	16104	10.037	ug/l	96
20) Methylene Chloride	2.788	84	4140	0.598	ug/l	94
30) Chloroform	5.092	83	7860	1.425	ug/l	99
47) Bromodichloromethane	7.817	83	3207	0.885	ug/l #	88
60) Dibromochloromethane	9.524	129	1358	0.555	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053023\
Data File : VX035916.D
Acq On : 30 May 2023 19:08
Operator : JC/MD
Sample : 02924-01RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001-002-003-1RE

Quant Time: May 31 05:49:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X052623W.M
Quant Title : SW846 8260
QLast Update : Mon May 29 07:45:36 2023
Response via : Initial Calibration

