

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053024\
 Data File : VX041688.D
 Acq On : 30 May 2024 12:33
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: May 31 05:29:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 16 05:13:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	24417	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	126731	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	115814	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	58272	30.305	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.033%
60) 4-Bromofluorobenzene	11.079	95	49884	27.592	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.967%
63) Toluene-d8	8.647	98	143905	28.987	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.633%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053024\
Data File : VX041688.D
Acq On : 30 May 2024 12:33
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: May 31 05:29:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 16 05:13:13 2024
Response via : Initial Calibration

