

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053024\
 Data File : VX041688.D
 Acq On : 30 May 2024 12:33
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: May 31 05:29:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu May 16 05:13:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.897	128	24417	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.757	114	126731	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	115814	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	58272	30.305	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.033%
60) 4-Bromofluorobenzene	11.079	95	49884	27.592	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.967%
63) Toluene-d8	8.647	98	143905	28.987	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.633%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053024\
Data File : VX041688.D
Acq On : 30 May 2024 12:33
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: May 31 05:29:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X051524W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu May 16 05:13:13 2024
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 4.897 min Scan# 61
Delta R.T. 0.000 min
Lab File: VX041688.D
Acq: 30 May 2024 12:33




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:128 Resp: 24417
Ion Ratio Lower Upper
128 100
49 151.3 0.0 361.5
130 132.6 0.0 324.3



#27
1,2-Dichloroethane-d4
Concen: 30.305 ug/l
RT: 5.952 min Scan# 798
Delta R.T. -0.000 min
Lab File: VX041688.D
Acq: 30 May 2024 12:33

Tgt Ion: 65 Resp: 58272
Ion Ratio Lower Upper
65 100
67 50.9 40.8 61.2









#28
1,4-Difluorobenzene
Concen: 30.000 ug/l
RT: 6.757 min Scan# 91
Delta R.T. -0.006 min
Lab File: VX041688.D
Acq: 30 May 2024 12:33




Instrument :
MSVOA_X
ClientSampleId :
IBLK


Tgt Ion:114 Resp: 126731
Ion Ratio Lower Upper
114 100
63 20.0 15.4 23.0
88 15.4 12.2 18.2



#57
Chlorobenzene-d5
Concen: 30.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX041688.D
Acq: 30 May 2024 12:33

Tgt Ion:117 Resp: 115814
Ion Ratio Lower Upper
117 100
82 51.6 41.7 62.5
119 32.6 25.7 38.5









#60
4-Bromofluorobenzene
Concen: 27.592 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX041688.D
Acq: 30 May 2024 12:33

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Tgt Ion: 95 Resp: 49884
Ion Ratio Lower Upper
95 100
174 95.6 76.6 115.0
176 94.4 75.4 113.0



#63
Toluene-d8
Concen: 28.987 ug/l
RT: 8.647 min Scan# 1240
Delta R.T. -0.000 min
Lab File: VX041688.D
Acq: 30 May 2024 12:33

Tgt Ion: 98 Resp: 143905
Ion Ratio Lower Upper
98 100
100 65.4 53.1 79.7

