

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046414.D
 Acq On : 30 May 2025 12:18
 Operator : JC/MD
 Sample : PB168191TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB168191TB

Quant Time: May 30 23:25:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	59922	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	119336	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	113811	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	48567	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60371	54.041	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 108.080%		
35) Dibromofluoromethane	5.379	113	43701	50.854	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.700%		
50) Toluene-d8	8.647	98	154072	51.801	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 103.600%		
62) 4-Bromofluorobenzene	11.079	95	58549	51.318	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 102.640%		
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	30025	67.032	ug/l	99
18) Methyl Acetate	2.709	43	1623	1.561	ug/l	93
20) Methylene Chloride	2.788	84	4493	5.234	ug/l #	95
43) Isopropyl Acetate	6.342	43	11813	5.428	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046414.D
Acq On : 30 May 2025 12:18
Operator : JC/MD
Sample : PB168191TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB168191TB

Quant Time: May 30 23:25:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

