

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046417.D
 Acq On : 30 May 2025 13:29
 Operator : JC/MD
 Sample : Q2151-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-1

Quant Time: May 30 23:27:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	58962	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	121344	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	115697	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	47824	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60543	55.077	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.160%	
35) Dibromofluoromethane	5.379	113	44027	50.385	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.780%	
50) Toluene-d8	8.647	98	154738	51.164	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.320%	
62) 4-Bromofluorobenzene	11.079	95	59889	51.624	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 103.240%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	20104	45.614	ug/l	92
18) Methyl Acetate	2.703	43	13811	13.504	ug/l	99
20) Methylene Chloride	2.782	84	2630	3.114	ug/l	88
37) Ethyl Acetate	4.733	43	7950	5.481	ug/l	95
43) Isopropyl Acetate	6.342	43	5978	2.701	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046417.D
Acq On : 30 May 2025 13:29
Operator : JC/MD
Sample : Q2151-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-1

Quant Time: May 30 23:27:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

