

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046418.D
 Acq On : 30 May 2025 13:52
 Operator : JC/MD
 Sample : Q2149-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FILTER-CAKE

Quant Time: May 30 23:27:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62556	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	124459	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	115572	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	52207	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64010	54.886	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.780%	
35) Dibromofluoromethane	5.379	113	46528	51.915	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.820%	
50) Toluene-d8	8.647	98	158494	51.094	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 102.180%	
62) 4-Bromofluorobenzene	11.079	95	61081	51.334	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 102.660%	
Target Compounds						
16) Acetone	2.386	43	2562	5.479	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046418.D
Acq On : 30 May 2025 13:52
Operator : JC/MD
Sample : Q2149-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FILTER-CAKE

Quant Time: May 30 23:27:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

