

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
 Data File : VX046423.D
 Acq On : 30 May 2025 15:49
 Operator : JC/MD
 Sample : Q2162-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-182-GW-620-622

Quant Time: May 30 23:30:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
 Quant Title : SW846 8260
 QLast Update : Tue May 06 07:12:22 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	60158	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	121355	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111155	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	45099	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	60632	54.061	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.120%	
35) Dibromofluoromethane	5.385	113	44162	50.535	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 101.080%	
50) Toluene-d8	8.647	98	150641	49.805	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 99.600%	
62) 4-Bromofluorobenzene	11.079	95	56303	48.529	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	= 97.060%	
Target Compounds						
16) Acetone	2.380	43	5821	12.945	ug/l	Qvalue # 86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053025\
Data File : VX046423.D
Acq On : 30 May 2025 15:49
Operator : JC/MD
Sample : Q2162-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-182-GW-620-622

Quant Time: May 30 23:30:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X050525W.M
Quant Title : SW846 8260
QLast Update : Tue May 06 07:12:22 2025
Response via : Initial Calibration

