

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\
 Data File : VX009987.D
 Acq On : 31 May 2019 16:00
 Operator : JC/SP
 Sample : K3131-09
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014MW028-190530

Quant Time: Jun 01 05:21:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	183325	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	295653	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	271447	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	123449	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	123571	47.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.44%	
35) Dibromofluoromethane	5.49	113	91365	49.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.30%	
50) Toluene-d8	8.71	98	376226	49.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.68%	
62) 4-Bromofluorobenzene	11.13	95	132151	48.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.48%	
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.41	62	4724	1.740	ug/l #	81
27) cis-1,2-Dichloroethene	4.60	96	1265	0.530	ug/l	73
40) Benzene	6.14	78	2513	0.282	ug/l #	75
52) Toluene	8.79	92	2292	0.424	ug/l	93
65) Chlorobenzene	10.14	112	1572922	276.788	ug/l	99
91) 1,2-Dichlorobenzene	12.39	146	6112	1.500	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX053119\
Data File : VX009987.D
Acq On : 31 May 2019 16:00
Operator : JC/SP
Sample : K3131-09
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014MW028-190530

Quant Time: Jun 01 05:21:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

