

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\
 Data File : VX009990.D
 Acq On : 31 May 2019 17:10
 Operator : JC/SP
 Sample : K3131-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF014MW026-190530

Quant Time: Jun 01 05:26:15 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	179226	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	292816	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	262663	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	116598	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	123264	48.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.38%	
35) Dibromofluoromethane	5.50	113	89850	48.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.62%	
50) Toluene-d8	8.71	98	370721	49.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.18%	
62) 4-Bromofluorobenzene	11.13	95	128758	47.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.90%	
Target Compounds						
44) Trichloroethene	7.21	130	4167	1.947	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX053119\
Data File : VX009990.D
Acq On : 31 May 2019 17:10
Operator : JC/SP
Sample : K3131-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF014MW026-190530

Quant Time: Jun 01 05:26:15 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

