

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\
 Data File : VX009993.D
 Acq On : 31 May 2019 18:20
 Operator : JC/SP
 Sample : K3131-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z1-TB05-190530

Quant Time: Jun 01 05:33:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	183882	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	296748	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	260763	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	108985	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	127622	49.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.26%	
35) Dibromofluoromethane	5.50	113	92941	49.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.64%	
50) Toluene-d8	8.71	98	374811	49.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.94%	
62) 4-Bromofluorobenzene	11.13	95	123411	45.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX053119\
Data File : VX009993.D
Acq On : 31 May 2019 18:20
Operator : JC/SP
Sample : K3131-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z1-TB05-190530

Quant Time: Jun 01 05:33:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

