

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\
 Data File : VX010000.D
 Acq On : 31 May 2019 22:14
 Operator : JC/SP
 Sample : VX0531WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0531WBL02

Quant Time: Jun 01 06:25:31 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	176231	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	282245	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	254554	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	114999	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	125028	50.22	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.44%	
35) Dibromofluoromethane	5.50	113	92241	51.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.96%	
50) Toluene-d8	8.71	98	377167	52.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.68%	
62) 4-Bromofluorobenzene	11.13	95	129432	50.01	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.02%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX053119\
Data File : VX010000.D
Acq On : 31 May 2019 22:14
Operator : JC/SP
Sample : VX0531WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0531WBL02

Quant Time: Jun 01 06:25:31 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

