

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\
 Data File : VX010003.D
 Acq On : 31 May 2019 23:24
 Operator : JC/SP
 Sample : K3128-01DL 5X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S32-0659DL

Quant Time: Jun 01 06:32:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	179623	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	286332	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	260408	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	117894	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	117786	46.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.84%	
35) Dibromofluoromethane	5.50	113	87463	48.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.18%	
50) Toluene-d8	8.71	98	363053	49.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.32%	
62) 4-Bromofluorobenzene	11.13	95	124323	47.35	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.70%	
Target Compounds						
4) Vinyl Chloride	1.41	62	2437	0.916	ug/l	91
21) trans-1,2-Dichloroethene	3.17	96	11635	5.748	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	83305	35.655	ug/l	88
44) Trichloroethene	7.22	130	2717	1.298	ug/l	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX053119\
Data File : VX010003.D
Acq On : 31 May 2019 23:24
Operator : JC/SP
Sample : K3128-01DL 5X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S32-0659DL

Quant Time: Jun 01 06:32:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

