

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\  
 Data File : VX010006.D  
 Acq On : 01 Jun 2019 00:34  
 Operator : JC/SP  
 Sample : K3128-03  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 38 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 S32-0661

Quant Time: Jun 01 06:39:32 2019  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X052819W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed May 29 04:00:56 2019  
 Response via : Initial Calibration

| Internal Standards           | R.T.  | QIon | Response | Conc   | Units  | Dev(Min) |
|------------------------------|-------|------|----------|--------|--------|----------|
| 1) Pentafluorobenzene        | 5.67  | 168  | 178568   | 50.00  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene      | 6.86  | 114  | 283425   | 50.00  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5         | 10.11 | 117  | 258065   | 50.00  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4   | 12.08 | 152  | 116493   | 50.00  | ug/l   | 0.00     |
| System Monitoring Compounds  |       |      |          |        |        |          |
| 33) 1,2-Dichloroethane-d4    | 6.06  | 65   | 118997   | 47.18  | ug/l   | 0.00     |
| Spiked Amount 50.000         |       |      | Recovery | =      | 94.36% |          |
| 35) Dibromofluoromethane     | 5.50  | 113  | 88357    | 49.59  | ug/l   | 0.00     |
| Spiked Amount 50.000         |       |      | Recovery | =      | 99.18% |          |
| 50) Toluene-d8               | 8.71  | 98   | 360648   | 49.84  | ug/l   | 0.00     |
| Spiked Amount 50.000         |       |      | Recovery | =      | 99.68% |          |
| 62) 4-Bromofluorobenzene     | 11.13 | 95   | 124613   | 47.95  | ug/l   | 0.00     |
| Spiked Amount 50.000         |       |      | Recovery | =      | 95.90% |          |
| Target Compounds             |       |      |          |        |        |          |
|                              |       |      |          |        | Qvalue |          |
| 2) Dichlorodifluoromethane   | 1.20  | 85   | 1583     | 0.725  | ug/l   | 94       |
| 4) Vinyl Chloride            | 1.41  | 62   | 1158     | 0.438  | ug/l   | 94       |
| 16) Acetone                  | 2.44  | 43   | 4369     | 2.998  | ug/l   | 96       |
| 21) trans-1,2-Dichloroethene | 3.17  | 96   | 1400     | 0.696  | ug/l   | 93       |
| 27) cis-1,2-Dichloroethene   | 4.60  | 96   | 41206    | 17.741 | ug/l   | 86       |
| 44) Trichloroethene          | 7.21  | 130  | 11025    | 5.321  | ug/l   | 94       |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX053119\  
Data File : VX010006.D  
Acq On : 01 Jun 2019 00:34  
Operator : JC/SP  
Sample : K3128-03  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 38 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
S32-0661

Quant Time: Jun 01 06:39:32 2019  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X052819W.M  
Quant Title : SW846 8260  
QLast Update : Wed May 29 04:00:56 2019  
Response via : Initial Calibration

