

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\
 Data File : VX010007.D
 Acq On : 01 Jun 2019 00:58
 Operator : JC/SP
 Sample : K3093-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 13908

Quant Time: Jun 01 06:42:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	175928	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	280908	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	253595	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	114834	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	116891	47.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.08%	
35) Dibromofluoromethane	5.50	113	86278	48.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.70%	
50) Toluene-d8	8.71	98	354946	49.49	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.98%	
62) 4-Bromofluorobenzene	11.13	95	126498	49.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.22%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.04	59	7193	11.761	ug/l #	86
16) Acetone	2.44	43	341742	238.006	ug/l	99
25) 2-Butanone	4.67	43	259408	115.960	ug/l	98
37) Ethyl Acetate	4.84	43	156577	39.427	ug/l	99
51) 4-Methyl-2-Pentanone	8.64	43	25194	6.044	ug/l	100
52) Toluene	8.79	92	11643	2.268	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX053119\
Data File : VX010007.D
Acq On : 01 Jun 2019 00:58
Operator : JC/SP
Sample : K3093-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
13908

Quant Time: Jun 01 06:42:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

