

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\
 Data File : VX010009.D
 Acq On : 01 Jun 2019 01:45
 Operator : JC/SP
 Sample : K3129-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW003-190528

Quant Time: Jun 01 06:46:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	179425	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	283611	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	256131	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	114635	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	116737	46.06	ug/l	0.00
Spiked Amount			Recovery	=	92.12%	
35) Dibromofluoromethane	5.50	113	88292	49.52	ug/l	0.00
Spiked Amount			Recovery	=	99.04%	
50) Toluene-d8	8.71	98	358652	49.53	ug/l	0.00
Spiked Amount			Recovery	=	99.06%	
62) 4-Bromofluorobenzene	11.13	95	123837	47.62	ug/l	0.00
Spiked Amount			Recovery	=	95.24%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.41	62	46623	17.546	ug/l	100
12) 1,1-Dichloroethene	2.37	96	2495	1.373	ug/l	93
21) trans-1,2-Dichloroethene	3.17	96	3905	1.931	ug/l	90
27) cis-1,2-Dichloroethene	4.60	96	2323354	995.503	ug/l	86
40) Benzene	6.15	78	3335	0.391	ug/l	91
65) Chlorobenzene	10.14	112	27607	5.149	ug/l	98
91) 1,2-Dichlorobenzene	12.39	146	7926	2.095	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX053119\
Data File : VX010009.D
Acq On : 01 Jun 2019 01:45
Operator : JC/SP
Sample : K3129-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW003-190528

Quant Time: Jun 01 06:46:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

