

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\
 Data File : VX010020.D
 Acq On : 01 Jun 2019 06:02
 Operator : JC/SP
 Sample : K3129-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF012MW004-190528

Quant Time: Jun 03 00:50:45 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	168889	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	269020	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	244656	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	109845	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	112282	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	
35) Dibromofluoromethane	5.50	113	82563	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
50) Toluene-d8	8.71	98	339589	49.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.88%	
62) 4-Bromofluorobenzene	11.13	95	117273	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	236405	107.613	ug/l	86
44) Trichloroethene	7.22	130	1220	0.620	ug/l	80
91) 1,2-Dichlorobenzene	12.39	146	1852	0.511	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\
Data File : VX010020.D
Acq On : 01 Jun 2019 06:02
Operator : JC/SP
Sample : K3129-15
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF012MW004-190528

Quant Time: Jun 03 00:50:45 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

