

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX053119\
 Data File : VX010023.D
 Acq On : 01 Jun 2019 07:12
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 03 00:59:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.611	0.637	-4.3	83	0.00
3 P	Chloromethane	0.821	0.700	14.7	78	0.00
4 C	Vinyl Chloride	0.740	0.663	10.4#	79	0.00
5 T	Bromomethane	0.449	0.364	18.9	78	0.00
6 T	Chloroethane	0.456	0.402	11.8	82	0.00
7 T	Trichlorofluoromethane	0.818	0.766	6.4	82	0.00
8 T	Diethyl Ether	0.385	0.335	13.0	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.494	0.458	7.3	83	0.00
10 T	Methyl Iodide	0.551	0.496	10.0	73	0.00
11 T	Tert butyl alcohol	0.174	0.186	-6.9	100	0.00
12 CM	1,1-Dichloroethene	0.506	0.476	5.9#	85	0.00
13 T	Acrolein	0.110	0.109	0.9	81	0.00
14 T	Allyl chloride	1.231	1.097	10.9	80	0.00
15 T	Acrylonitrile	0.419	0.393	6.2	88	0.00
16 T	Acetone	0.408	0.355	13.0	81	0.00
17 T	Carbon Disulfide	1.653	1.390	15.9	82	0.00
18 T	Methyl Acetate	0.950	0.960	-1.1	99	0.00
19 T	Methyl tert-butyl Ether	1.912	1.839	3.8	88	0.00
20 T	Methylene Chloride	0.625	0.580	7.2	88	0.00
21 T	trans-1,2-Dichloroethene	0.563	0.511	9.2	87	0.00
22 T	Diisopropyl ether	2.342	2.225	5.0	86	0.00
23 T	Vinyl Acetate	2.104	1.939	7.8	83	0.00
24 P	1,1-Dichloroethane	1.135	1.069	5.8	86	0.00
25 T	2-Butanone	0.636	0.601	5.5	87	0.00
26 T	2,2-Dichloropropane	0.858	0.480	44.1#	51	0.00
27 T	cis-1,2-Dichloroethene	0.650	0.605	6.9	89	0.00
28 T	Bromochloromethane	0.589	0.507	13.9	83	0.00
29 T	Tetrahydrofuran	0.408	0.394	3.4	89	0.00
30 C	Chloroform	1.023	0.979	4.3#	88	0.00
31 T	Cyclohexane	1.075	1.024	4.7	82	0.00
32 T	1,1,1-Trichloroethane	0.833	0.822	1.3	87	0.00
33 S	1,2-Dichloroethane-d4	0.706	0.692	2.0	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	90	0.00
35 S	Dibromofluoromethane	0.314	0.320	-1.9	95	0.00
36 T	1,1-Dichloropropene	0.507	0.468	7.7	86	0.00
37 T	Ethyl Acetate	0.707	0.696	1.6	88	0.00
38 T	Carbon Tetrachloride	0.438	0.434	0.9	88	0.00
39 T	Methylcyclohexane	0.602	0.564	6.3	82	0.00
40 TM	Benzene	1.505	1.438	4.5	87	0.00
41 T	Methacrylonitrile	0.413	0.396	4.1	89	0.00
42 TM	1,2-Dichloroethane	0.556	0.530	4.7	89	0.00
43 T	Isopropyl Acetate	1.141	1.089	4.6	88	0.00
44 TM	Trichloroethene	0.366	0.346	5.5	88	0.00
45 C	1,2-Dichloropropane	0.433	0.408	5.8#	87	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\
 Data File : VX010023.D
 Acq On : 01 Jun 2019 07:12
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 03 00:59:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 29 04:00:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.255	0.233	8.6	87	0.00
47 T	Bromodichloromethane	0.485	0.480	1.0	89	0.00
48 T	Methyl methacrylate	0.565	0.555	1.8	88	0.00
49 T	1,4-Dioxane	0.011	0.011	0.0	94	0.00
50 S	Toluene-d8	1.277	1.303	-2.0	93	0.00
51 T	4-Methyl-2-Pentanone	0.742	0.733	1.2	89	0.00
52 CM	Toluene	0.914	0.893	2.3#	88	0.00
53 T	t-1,3-Dichloropropene	0.582	0.525	9.8	80	0.00
54 T	cis-1,3-Dichloropropene	0.634	0.574	9.5	81	0.00
55 T	1,1,2-Trichloroethane	0.368	0.357	3.0	89	0.00
56 T	Ethyl methacrylate	0.655	0.666	-1.7	87	0.00
57 T	1,3-Dichloropropane	0.660	0.636	3.6	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.343	0.337	1.7	86	0.00
59 T	2-Hexanone	0.571	0.570	0.2	87	0.00
60 T	Dibromochloromethane	0.362	0.371	-2.5	90	0.00
61 T	1,2-Dibromoethane	0.378	0.365	3.4	89	0.00
62 S	4-Bromofluorobenzene	0.459	0.481	-4.8	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.349	0.374	-7.2	97	0.00
65 PM	Chlorobenzene	1.047	1.007	3.8	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.363	0.372	-2.5	91	0.00
67 C	Ethyl Benzene	1.879	1.856	1.2#	88	0.00
68 T	m/p-Xylenes	0.694	0.680	2.0	87	0.00
69 T	o-Xylene	0.666	0.663	0.5	88	0.00
70 T	Styrene	1.168	1.175	-0.6	89	0.00
71 P	Bromoform	0.297	0.317	-6.7	93	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
73 T	Isopropylbenzene	3.689	3.566	3.3	88	0.00
74 T	N-amyl acetate	2.204	2.154	2.3	87	0.00
75 P	1,1,2,2-Tetrachloroethane	1.376	1.306	5.1	89	0.00
76 T	1,2,3-Trichloropropane	1.196	1.129	5.6	87	0.00
77 T	Bromobenzene	0.940	0.914	2.8	92	0.00
78 T	n-propylbenzene	4.252	4.158	2.2	88	0.00
79 T	2-Chlorotoluene	2.598	2.472	4.8	88	0.00
80 T	1,3,5-Trimethylbenzene	3.108	3.047	2.0	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.483	0.393	18.6	73	0.00
82 T	4-Chlorotoluene	3.024	2.863	5.3	87	0.00
83 T	tert-Butylbenzene	2.973	2.893	2.7	88	0.00
84 T	1,2,4-Trimethylbenzene	3.156	3.061	3.0	88	0.00
85 T	sec-Butylbenzene	3.537	3.475	1.8	88	0.00
86 T	p-Isopropyltoluene	3.182	3.186	-0.1	87	0.00
87 T	1,3-Dichlorobenzene	1.674	1.599	4.5	90	0.00
88 T	1,4-Dichlorobenzene	1.707	1.589	6.9	89	0.00
89 T	n-Butylbenzene	2.990	2.830	5.4	83	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX053119\
Data File : VX010023.D
Acq On : 01 Jun 2019 07:12
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 03 00:59:00 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052819W.M
Quant Title : SW846 8260
QLast Update : Wed May 29 04:00:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.546	0.560	-2.6	91	0.00
91 T	1,2-Dichlorobenzene	1.650	1.594	3.4	90	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.324	0.304	6.2	89	0.00
93 T	1,2,4-Trichlorobenzene	1.179	1.135	3.7	89	0.00
94 T	Hexachlorobutadiene	0.592	0.558	5.7	86	0.00
95 T	Naphthalene	3.670	3.647	0.6	91	0.00
96 T	1,2,3-Trichlorobenzene	1.180	1.133	4.0	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6