

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053122\
 Data File : VX029068.D
 Acq On : 31 May 2022 11:44
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 01 04:56:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 01 04:45:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	342691	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	575067	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	557900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	269255	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	188159	48.691	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.380%
35) Dibromofluoromethane	5.385	113	192143	50.303	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	100.600%
50) Toluene-d8	8.653	98	675596	48.359	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	96.720%
62) 4-Bromofluorobenzene	11.079	95	260774	48.289	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053122\
Data File : VX029068.D
Acq On : 31 May 2022 11:44
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 01 04:56:29 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053122W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 01 04:45:42 2022
Response via : Initial Calibration

