

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053123\
 Data File : VX035936.D
 Acq On : 31 May 2023 15:11
 Operator : JC/MD
 Sample : 02955-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: May 31 17:08:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 31 12:50:30 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	217675	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	353942	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	312259	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	147640	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	142600	40.343	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	80.680%
35) Dibromofluoromethane	5.385	113	104477	43.887	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.780%
50) Toluene-d8	8.647	98	411181	48.126	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.260%
62) 4-Bromofluorobenzene	11.079	95	158242	46.918	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX053123\
Data File : VX035936.D
Acq On : 31 May 2023 15:11
Operator : JC/MD
Sample : 02955-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: May 31 17:08:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X053123W.M
Quant Title : SW846 8260
QLast Update : Wed May 31 12:50:30 2023
Response via : Initial Calibration

