

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060120\
 Data File : VX016519.D
 Acq On : 01 Jun 2020 11:34
 Operator : JC/SP
 Sample : L2799-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 Z2-TB7-200527

Quant Time: Jun 02 02:01:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	346679	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	556499	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	562794	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	282618	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	197967	45.98	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.96%	
35) Dibromofluoromethane	5.46	113	158482	46.25	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.50%	
50) Toluene-d8	8.70	98	698078	52.34	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.68%	
62) 4-Bromofluorobenzene	11.12	95	284461	55.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	110.40%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060120\
Data File : VX016519.D
Acq On : 01 Jun 2020 11:34
Operator : JC/SP
Sample : L2799-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
Z2-TB7-200527

Quant Time: Jun 02 02:01:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

