

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060120\
 Data File : VX016530.D
 Acq On : 01 Jun 2020 15:44
 Operator : JC/SP
 Sample : L2801-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW608-200528

Quant Time: Jun 02 02:24:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	262644	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	426029	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	424861	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	224954	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	155340	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
35) Dibromofluoromethane	5.47	113	123881	47.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.44%	
50) Toluene-d8	8.70	98	531962	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
62) 4-Bromofluorobenzene	11.12	95	220122	55.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.58%	
Target Compounds						
16) Acetone	2.42	43	2292	1.604	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	21516	6.653	ug/l	92
44) Trichloroethene	7.19	130	7942	2.084	ug/l	98
52) Toluene	8.77	92	1897	0.250	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060120\
Data File : VX016530.D
Acq On : 01 Jun 2020 15:44
Operator : JC/SP
Sample : L2801-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW608-200528

Quant Time: Jun 02 02:24:17 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

