

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060120\
 Data File : VX016533.D
 Acq On : 01 Jun 2020 16:52
 Operator : JC/SP
 Sample : L2802-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW113-200527

Quant Time: Jun 02 08:54:44 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	256349	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	417644	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	424399	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	222189	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	151763	47.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.34%	
35) Dibromofluoromethane	5.46	113	122101	47.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.96%	
50) Toluene-d8	8.70	98	526128	52.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.12%	
62) 4-Bromofluorobenzene	11.12	95	213836	55.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.58%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.31	50	1357	0.445	ug/l	100
4) Vinyl Chloride	1.39	62	3295	1.022	ug/l	91
16) Acetone	2.42	43	5099	3.656	ug/l	92
17) Carbon Disulfide	2.55	76	2447	0.334	ug/l #	78
27) cis-1,2-Dichloroethene	4.57	96	6093	1.930	ug/l	84
29) Tetrahydrofuran	5.11	42	3132	2.051	ug/l	91
52) Toluene	8.77	92	8688	1.170	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060120\
Data File : VX016533.D
Acq On : 01 Jun 2020 16:52
Operator : JC/SP
Sample : L2802-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW113-200527

Quant Time: Jun 02 08:54:44 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

