

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060120\
 Data File : VX016534.D
 Acq On : 01 Jun 2020 17:15
 Operator : JC/SP
 Sample : L2801-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS050MW638-200528

Quant Time: Jun 02 03:08:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	265286	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	427772	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	428837	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	228647	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	157117	47.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.38%	
35) Dibromofluoromethane	5.47	113	125669	47.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.42%	
50) Toluene-d8	8.70	98	535211	52.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.42%	
62) 4-Bromofluorobenzene	11.12	95	219661	55.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.90%	
Target Compounds						
3) Chloromethane	1.31	50	1813	0.575	ug/l	98
4) Vinyl Chloride	1.39	62	873	0.262	ug/l	94
16) Acetone	2.42	43	5632	3.902	ug/l	99
27) cis-1,2-Dichloroethene	4.57	96	3832	1.173	ug/l	86
40) Benzene	6.11	78	5331	0.442	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060120\
Data File : VX016534.D
Acq On : 01 Jun 2020 17:15
Operator : JC/SP
Sample : L2801-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS050MW638-200528

Quant Time: Jun 02 03:08:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

