

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX060120\
 Data File : VX016540.D
 Acq On : 01 Jun 2020 19:31
 Operator : JC/SP
 Sample : IBLK
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 02 03:36:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	254382	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	411946	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	418756	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	223893	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	151015	47.80	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.60%	
35) Dibromofluoromethane	5.46	113	121174	47.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.54%	
50) Toluene-d8	8.70	98	510726	51.73	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.46%	
62) 4-Bromofluorobenzene	11.12	95	216601	56.78	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.56%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	3.00	59	1830	2.259	ug/l #	72
16) Acetone	2.42	43	15901	11.489	ug/l	93
25) 2-Butanone	4.65	43	3501	1.493	ug/l	93
51) 4-Methyl-2-Pentanone	8.62	43	1383	0.304	ug/l	94
59) 2-Hexanone	9.48	43	1013	0.286	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX060120\
Data File : VX016540.D
Acq On : 01 Jun 2020 19:31
Operator : JC/SP
Sample : IBLK
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 02 03:36:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

