

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX060120\
 Data File : VX016516.D
 Acq On : 01 Jun 2020 09:51
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 02 01:53:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	0.492	0.514	-4.5	104	0.00
3 P	Chloromethane	0.595	0.564	5.2	101	0.00
4 C	Vinyl Chloride	0.629	0.639	-1.6#	106	0.00
5 T	Bromomethane	0.306	0.308	-0.7	110	0.00
6 T	Chloroethane	0.367	0.382	-4.1	107	0.00
7 T	Trichlorofluoromethane	0.826	0.866	-4.8	108	0.00
8 T	Diethyl Ether	0.344	0.345	-0.3	108	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.455	0.490	-7.7	112	0.00
10 T	Methyl Iodide	0.490	0.530	-8.2	106	0.00
11 T	Tert butyl alcohol	0.159	0.141	11.3	91	-0.01
12 CM	1,1-Dichloroethene	0.481	0.498	-3.5#	109	0.00
13 T	Acrolein	0.068	0.070	-2.9	115	0.00
14 T	Allyl chloride	0.885	0.881	0.5	103	0.00
15 T	Acrylonitrile	0.326	0.312	4.3	97	0.00
16 T	Acetone	0.272	0.259	4.8	97	0.00
17 T	Carbon Disulfide	1.429	1.434	-0.3	104	0.00
18 T	Methyl Acetate	0.698	0.679	2.7	99	0.00
19 T	Methyl tert-butyl Ether	1.728	1.788	-3.5	105	0.00
20 T	Methylene Chloride	0.563	0.549	2.5	103	0.00
21 T	trans-1,2-Dichloroethene	0.532	0.545	-2.4	107	0.00
22 T	Diisopropyl ether	1.700	1.730	-1.8	103	0.00
23 T	Vinyl Acetate	1.488	1.513	-1.7	100	0.00
24 P	1,1-Dichloroethane	0.959	0.970	-1.1	103	0.00
25 T	2-Butanone	0.461	0.414	10.2	90	0.00
26 T	2,2-Dichloropropane	0.785	0.897	-14.3	122	0.00
27 T	cis-1,2-Dichloroethene	0.616	0.615	0.2	105	0.00
28 T	Bromochloromethane	0.446	0.424	4.9	104	0.00
29 T	Tetrahydrofuran	0.298	0.283	5.0	94	-0.01
30 C	Chloroform	0.967	0.977	-1.0#	104	0.00
31 T	Cyclohexane	0.863	0.867	-0.5	105	0.00
32 T	1,1,1-Trichloroethane	0.875	0.894	-2.2	104	0.00
33 S	1,2-Dichloroethane-d4	0.621	0.582	6.3	100	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.308	0.303	1.6	103	0.00
36 T	1,1-Dichloropropene	0.473	0.500	-5.7	107	0.00
37 T	Ethyl Acetate	0.568	0.545	4.0	96	0.00
38 T	Carbon Tetrachloride	0.497	0.533	-7.2	107	0.00
39 T	Methylcyclohexane	0.514	0.571	-11.1	112	0.00
40 TM	Benzene	1.410	1.478	-4.8	105	0.00
41 T	Methacrylonitrile	0.312	0.300	3.8	99	0.00
42 TM	1,2-Dichloroethane	0.501	0.512	-2.2	102	0.00
43 T	Isopropyl Acetate	0.887	0.878	1.0	99	-0.01
44 TM	Trichloroethene	0.447	0.491	-9.8	114	0.00
45 C	1,2-Dichloropropane	0.359	0.374	-4.2#	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\WX060120\
 Data File : VX016516.D
 Acq On : 01 Jun 2020 09:51
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 02 01:53:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 27 16:31:05 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.241	0.247	-2.5	103	0.00
47 T	Bromodichloromethane	0.499	0.526	-5.4	105	0.00
48 T	Methyl methacrylate	0.424	0.424	0.0	98	0.00
49 T	1,4-Dioxane	0.010	0.009	10.0	93	0.00
50 S	Toluene-d8	1.198	1.210	-1.0	104	0.00
51 T	4-Methyl-2-Pentanone	0.552	0.548	0.7	96	0.00
52 CM	Toluene	0.889	0.956	-7.5#	107	0.00
53 T	t-1,3-Dichloropropene	0.568	0.622	-9.5	108	0.00
54 T	cis-1,3-Dichloropropene	0.601	0.658	-9.5	109	0.00
55 T	1,1,2-Trichloroethane	0.353	0.369	-4.5	105	0.00
56 T	Ethyl methacrylate	0.576	0.618	-7.3	104	0.00
57 T	1,3-Dichloropropane	0.607	0.640	-5.4	105	0.00
58 T	2-Chloroethyl Vinyl ether	0.300	0.317	-5.7	106	0.00
59 T	2-Hexanone	0.429	0.435	-1.4	97	0.00
60 T	Dibromochloromethane	0.398	0.426	-7.0	106	0.00
61 T	1,2-Dibromoethane	0.379	0.397	-4.7	105	0.00
62 S	4-Bromofluorobenzene	0.463	0.464	-0.2	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.559	0.625	-11.8	115	0.00
65 PM	Chlorobenzene	1.019	1.093	-7.3	107	0.00
66 T	1,1,1,2-Tetrachloroethane	0.389	0.414	-6.4	107	0.00
67 C	Ethyl Benzene	1.815	1.961	-8.0#	107	0.00
68 T	m/p-Xylenes	0.673	0.744	-10.5	108	0.00
69 T	o-Xylene	0.652	0.703	-7.8	107	0.00
70 T	Styrene	1.120	1.235	-10.3	106	0.00
71 P	Bromoform	0.329	0.357	-8.5	105	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
73 T	Isopropylbenzene	3.464	3.795	-9.6	107	0.00
74 T	N-amyl acetate	1.594	1.652	-3.6	100	0.00
75 P	1,1,2,2-Tetrachloroethane	0.868	0.819	5.6	90	0.00
76 T	1,2,3-Trichloropropane	1.099	1.157	-5.3	104	0.00
77 T	Bromobenzene	0.909	0.977	-7.5	108	0.00
78 T	n-propylbenzene	3.870	4.282	-10.6	107	0.00
79 T	2-Chlorotoluene	2.363	2.592	-9.7	108	0.00
80 T	1,3,5-Trimethylbenzene	2.893	3.229	-11.6	108	0.00
81 T	trans-1,4-Dichloro-2-butene	0.447	0.480	-7.4	107	0.00
82 T	4-Chlorotoluene	2.790	3.057	-9.6	107	0.00
83 T	tert-Butylbenzene	2.472	2.703	-9.3	105	0.00
84 T	1,2,4-Trimethylbenzene	2.925	3.236	-10.6	107	0.00
85 T	sec-Butylbenzene	3.101	3.462	-11.6	109	0.00
86 T	p-Isopropyltoluene	2.905	3.298	-13.5	109	0.00
87 T	1,3-Dichlorobenzene	1.574	1.688	-7.2	106	0.00
88 T	1,4-Dichlorobenzene	1.610	1.751	-8.8	109	0.00
89 T	n-Butylbenzene	2.419	2.843	-17.5	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX060120\
Data File : VX016516.D
Acq On : 01 Jun 2020 09:51
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 02 01:53:33 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X052720W.M
Quant Title : SW846 8260
QLast Update : Wed May 27 16:31:05 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.502	0.550	-9.6	106	0.00
91 T	1,2-Dichlorobenzene	1.525	1.621	-6.3	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.281	0.280	0.4	98	0.00
93 T	1,2,4-Trichlorobenzene	1.013	1.145	-13.0	111	0.00
94 T	Hexachlorobutadiene	0.398	0.506	-27.1#	131	0.00
95 T	Naphthalene	3.472	3.760	-8.3	104	0.00
96 T	1,2,3-Trichlorobenzene	0.992	1.119	-12.8	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6