

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022288.D
 Acq On : 01 Jun 2021 13:27
 Operator : JC/MD
 Sample : VX0601WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0601WBL01

Quant Time: Jun 01 16:07:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	70068	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	147183	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	128665	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49035	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	58901	54.50	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.00%
35) Dibromofluoromethane	5.397	113	45786	49.05	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.10%
50) Toluene-d8	8.653	98	180535	50.21	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.42%
62) 4-Bromofluorobenzene	11.085	95	62342	46.91	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.82%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022288.D
Acq On : 01 Jun 2021 13:27
Operator : JC/MD
Sample : VX0601WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0601WBL01

Quant Time: Jun 01 16:07:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

