

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022296.D
 Acq On : 01 Jun 2021 17:23
 Operator : JC/MD
 Sample : M2532-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-N-(MH-10.29)

Quant Time: Jun 02 03:22:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	65428	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	141383	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	127778	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	54444	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56283	55.77	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.54%
35) Dibromofluoromethane	5.397	113	42558	47.47	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.94%
50) Toluene-d8	8.653	98	173529	50.24	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.48%
62) 4-Bromofluorobenzene	11.085	95	63364	49.64	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.28%
Target Compounds						
16) Acetone	2.386	43	3788	8.62	ug/l #	82
52) Toluene	8.726	92	1044	0.41	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022296.D
Acq On : 01 Jun 2021 17:23
Operator : JC/MD
Sample : M2532-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-N-(MH-10.29)

Quant Time: Jun 02 03:22:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

