

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022297.D
 Acq On : 01 Jun 2021 17:47
 Operator : JC/MD
 Sample : M2569-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 C0AA1

Quant Time: Jun 02 03:22:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	67455	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	143095	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	123898	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	49663	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	57695	55.45	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.90%
35) Dibromofluoromethane	5.397	113	43714	48.17	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.34%
50) Toluene-d8	8.653	98	173087	49.52	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.04%
62) 4-Bromofluorobenzene	11.085	95	59947	46.40	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022297.D
Acq On : 01 Jun 2021 17:47
Operator : JC/MD
Sample : M2569-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
C0AA1

Quant Time: Jun 02 03:22:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

