

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022299.D
 Acq On : 01 Jun 2021 18:34
 Operator : JC/MD
 Sample : M2545-02DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 04BR-20210526DL

Quant Time: Jun 02 03:23:31 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	65646	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	136791	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118042	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	44876	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54616	53.93	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	107.86%
35) Dibromofluoromethane	5.397	113	42269	48.73	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.46%
50) Toluene-d8	8.653	98	169006	50.58	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.16%
62) 4-Bromofluorobenzene	11.085	95	57317	46.41	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.82%
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.173	85	1203	1.51	ug/l	94
4) Vinyl Chloride	1.374	62	4881	5.36	ug/l #	86
9) 1,1,2-Trichlorotrifluo...	2.331	101	9935	12.77	ug/l	91
12) 1,1-Dichloroethene	2.331	96	743	0.99	ug/l #	69
21) trans-1,2-Dichloroethene	3.093	96	931	1.14	ug/l	97
27) cis-1,2-Dichloroethene	4.495	96	50538	50.78	ug/l	90
44) Trichloroethene	7.129	130	76506	79.11	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022299.D
Acq On : 01 Jun 2021 18:34
Operator : JC/MD
Sample : M2545-02DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
04BR-20210526DL

Quant Time: Jun 02 03:23:31 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

