

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022305.D
 Acq On : 01 Jun 2021 20:55
 Operator : JC/MD
 Sample : M2560-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S1-A-BOTTOM

Quant Time: Jun 02 03:25:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	63947	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	136185	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121838	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48587	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53961	54.70	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.40%	
35) Dibromofluoromethane	5.397	113	40235	46.59	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 93.18%	
50) Toluene-d8	8.653	98	169412	50.92	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.84%	
62) 4-Bromofluorobenzene	11.085	95	59013	47.99	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.98%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	43602	101.51	ug/l	98
18) Methyl Acetate	2.715	43	1404	1.23	ug/l	90
20) Methylene Chloride	2.794	84	9066	9.51	ug/l	92
25) 2-Butanone	4.574	43	6016	8.66	ug/l	# 83
43) Isopropyl Acetate	6.348	43	12773	5.06	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022305.D
Acq On : 01 Jun 2021 20:55
Operator : JC/MD
Sample : M2560-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S1-A-BOTTOM

Quant Time: Jun 02 03:25:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

