

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022314.D
 Acq On : 02 Jun 2021 01:13
 Operator : JC/MD
 Sample : PB136731TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136731TB

Quant Time: Jun 02 05:32:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	67854	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	144300	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	130237	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	52672	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56865	54.33	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.66%	
35) Dibromofluoromethane	5.397	113	43968	48.05	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.10%	
50) Toluene-d8	8.653	98	178068	50.52	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.04%	
62) 4-Bromofluorobenzene	11.085	95	62751	48.16	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 96.32%	
Target Compounds						
						Qvalue
16) Acetone	2.386	43	13771	30.22	ug/l	95
20) Methylene Chloride	2.794	84	4494	4.44	ug/l	87
43) Isopropyl Acetate	6.348	43	14736	5.51	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022314.D
Acq On : 02 Jun 2021 01:13
Operator : JC/MD
Sample : PB136731TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136731TB

Quant Time: Jun 02 05:32:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

