

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022315.D
 Acq On : 02 Jun 2021 01:36
 Operator : JC/MD
 Sample : PB136731ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136731ZHE#01

Quant Time: Jun 02 05:33:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	64556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	137260	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.061	117	122380	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48208	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	56203	56.44	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 112.88%			
35) Dibromofluoromethane	5.391	113	41689	47.89	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 95.78%			
50) Toluene-d8	8.653	98	171531	51.16	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 102.32%			
62) 4-Bromofluorobenzene	11.085	95	59941	48.37	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 96.74%			
Target Compounds					Qvalue	
16) Acetone	2.386	43	13047	30.09	ug/l	95
20) Methylene Chloride	2.794	84	4588	4.77	ug/l #	77
43) Isopropyl Acetate	6.355	43	15045	5.91	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022315.D
Acq On : 02 Jun 2021 01:36
Operator : JC/MD
Sample : PB136731ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136731ZHE#01

Quant Time: Jun 02 05:33:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

