

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022316.D
 Acq On : 02 Jun 2021 02:00
 Operator : JC/MD
 Sample : PB136731ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB136731ZHE#02

Quant Time: Jun 02 05:33:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64195	50.00	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	135492	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	120211	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46624	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54372	54.91	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 109.82%		
35) Dibromofluoromethane	5.397	113	41340	48.11	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 96.22%		
50) Toluene-d8	8.653	98	169303	51.15	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 102.30%		
62) 4-Bromofluorobenzene	11.085	95	57235	46.78	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 93.56%		
Target Compounds						
16) Acetone	2.386	43	13391	31.06	ug/l	Qvalue 91
18) Methyl Acetate	2.715	43	1648	1.44	ug/l	# 88
20) Methylene Chloride	2.794	84	4587	4.79	ug/l	# 81
43) Isopropyl Acetate	6.354	43	15071	6.00	ug/l	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022316.D
Acq On : 02 Jun 2021 02:00
Operator : JC/MD
Sample : PB136731ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB136731ZHE#02

Quant Time: Jun 02 05:33:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

