

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022317.D
 Acq On : 02 Jun 2021 02:23
 Operator : JC/MD
 Sample : M2572-01 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 25BR-20210527

Quant Time: Jun 02 05:33:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	64538	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	136479	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118279	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	45314	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54542	54.79	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	109.58%		
35) Dibromofluoromethane	5.398	113	41097	47.48	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	94.96%		
50) Toluene-d8	8.653	98	166205	49.85	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.70%		
62) 4-Bromofluorobenzene	11.086	95	57723	46.84	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	93.68%		
Target Compounds						Qvalue
2) Dichlorodifluoromethane	1.173	85	997	1.27	ug/l	83
4) Vinyl Chloride	1.374	62	27549	30.76	ug/l	98
12) 1,1-Dichloroethene	2.325	96	776	1.05	ug/l #	80
21) trans-1,2-Dichloroethene	3.093	96	4326	5.38	ug/l #	82
27) cis-1,2-Dichloroethene	4.495	96	394246	402.93	ug/l	91
44) Trichloroethene	7.129	130	158598	164.36	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022317.D
Acq On : 02 Jun 2021 02:23
Operator : JC/MD
Sample : M2572-01 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
25BR-20210527

Quant Time: Jun 02 05:33:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

