

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022318.D
 Acq On : 02 Jun 2021 02:47
 Operator : JC/MD
 Sample : M2572-02 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 45BR-20210528

Quant Time: Jun 02 05:34:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	63605	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	135970	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	121310	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	48073	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	54177	55.22	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.44%
35) Dibromofluoromethane	5.391	113	41197	47.78	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.56%
50) Toluene-d8	8.653	98	167377	50.39	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.78%
62) 4-Bromofluorobenzene	11.085	95	57894	47.16	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.32%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.374	62	2073	2.35	ug/l	100
12) 1,1-Dichloroethene	2.319	96	308	0.42	ug/l #	56
21) trans-1,2-Dichloroethene	3.087	96	403	0.51	ug/l #	72
27) cis-1,2-Dichloroethene	4.495	96	72550	75.24	ug/l	90
44) Trichloroethene	7.129	130	197891	205.85	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022318.D
Acq On : 02 Jun 2021 02:47
Operator : JC/MD
Sample : M2572-02 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
45BR-20210528

Quant Time: Jun 02 05:34:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

