

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022321.D
 Acq On : 02 Jun 2021 03:58
 Operator : JC/MD
 Sample : M2572-05 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20BR-20210528

Quant Time: Jun 02 05:35:01 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	64734	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	133662	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	118654	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46974	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	53904	53.98	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.96%	
35) Dibromofluoromethane	5.397	113	41112	48.50	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 97.00%	
50) Toluene-d8	8.653	98	165366	50.65	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.30%	
62) 4-Bromofluorobenzene	11.085	95	58744	48.68	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 97.36%	
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.173	85	1454	1.85	ug/l	97
4) Vinyl Chloride	1.374	62	39982	44.51	ug/l	100
9) 1,1,2-Trichlorotrifluo...	2.331	101	645	0.84	ug/l #	87
12) 1,1-Dichloroethene	2.325	96	757	1.02	ug/l	93
21) trans-1,2-Dichloroethene	3.105	96	930	1.15	ug/l	91
27) cis-1,2-Dichloroethene	4.495	96	270589	275.71	ug/l	90
44) Trichloroethene	7.129	130	67212	71.12	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022321.D
Acq On : 02 Jun 2021 03:58
Operator : JC/MD
Sample : M2572-05 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20BR-20210528

Quant Time: Jun 02 05:35:01 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

