

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
 Data File : VX022322.D
 Acq On : 02 Jun 2021 04:21
 Operator : JC/MD
 Sample : M2572-10 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 29BR-20210528

Quant Time: Jun 02 05:35:20 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
 Quant Title : SW846 8260
 QLast Update : Wed May 19 14:23:10 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	63377	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	136856	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	119592	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	46026	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	55083	56.34	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	112.68%
35) Dibromofluoromethane	5.391	113	41213	47.49	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.98%
50) Toluene-d8	8.653	98	165938	49.64	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.28%
62) 4-Bromofluorobenzene	11.085	95	57278	46.35	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.70%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.173	85	1992	2.59	ug/l	100
4) Vinyl Chloride	1.374	62	614	0.70	ug/l	99
9) 1,1,2-Trichlorotrifluo...	2.325	101	1700	2.26	ug/l	88
12) 1,1-Dichloroethene	2.319	96	1006	1.38	ug/l	84
21) trans-1,2-Dichloroethene	3.093	96	636	0.80	ug/l #	79
27) cis-1,2-Dichloroethene	4.495	96	19489	20.28	ug/l	92
44) Trichloroethene	7.129	130	197254	203.86	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX060121\
Data File : VX022322.D
Acq On : 02 Jun 2021 04:21
Operator : JC/MD
Sample : M2572-10 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
29BR-20210528

Quant Time: Jun 02 05:35:20 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X051921W.M
Quant Title : SW846 8260
QLast Update : Wed May 19 14:23:10 2021
Response via : Initial Calibration

